

DEVELOPMENT OF NEW METHODOLOGY FOR DOSE CALCULATION IN PHOTOGRAPHIC DOSIMETRY

Teresinha F. L. Daltro, Leticia L. Campos, Homero E.B. Perez

Instituto de Pesquisas Energéticas e Nucleares
Comissão Nacional de Energia Nuclear
Caixa Postal 11049, Pinheiros
CEP 05499-970, São Paulo -BRAZIL

INTRODUCTION

The personal dosimeter system of IPEN is based on film dosimetry. Personal doses at IPEN are mainly due to X or gamma radiation.

The use of personal photographic dosimeters involves two steps: firstly, data acquisition including their evaluation with respect to the calibration quantity and secondly, the interpretation of the data in terms of effective dose.

The effective dose was calculated using artificial intelligence techniques by means of neural network. The learning of the neural network was performed by taking the readings of optical density as a function of incident energy and exposure from the calibration curve (1). The obtained output in the daily grind is the mean effective energy and the effective dose.

MATERIALS AND METHODS

The photographic dosimeters used at IPEN are made up of two parts:

a) the Personal Monitoring 2/10 film (Agfa - Gevaert), consisting of two emulsions. b) the holder (badge) containing four filters (plastic, lead, copper and cadmium) and an open window. The area of the filters is sufficiently large to avoid edge effects.

The irradiations were performed using a ^{60}Co (1.0 GBq), ^{137}Cs (53 TBq) and a X-ray machine Stabilipan model 250 Siemens. The quality control was carried out using a ^{14}C plane source.

The measurement of radiographic film density was performed using a transmission densitometer MacBeth TD 904.

The calibration curves (OD x Exposure) were plotted out as a function of the radiation energy for each filter of the badge. In this case, for each lecture j , due to a certain filter, it is possible to construct a family of response curves of OD (d_j) as a function of Exposure (X_i) for each energy value (2).

NEURAL NETWORK STRUCTURE

Different structures of neural networks can be used, each one associated with one application. In this work it was used the backpropagation paradigma, where the neuronios are arranged in layers: input, output and intermediate layers as shown in Fig. 1 In this work it was used the Neural Works Professional II/Plus software (3,4).

All values used for the network training are the average of five lectures.

For the training of the network the data were arranged as a matrix as shown bellow. The matrix was presented to the network 10^6 times, during 30h in a PC-AT 486. The obtained average square error was less than 0.1%.

Different from the training, the use of the trained network is very fast, only a fraction of seconds (5).

In order to apply the trained network it was developed a software in Visual Basic Language for Windows.

$$\begin{bmatrix} DO - Cd_i & DO - Ja_i & DO - Cu_i & DO - Pb_i & DO - Pl_i & E_i & D_i \\ \cdot & \cdot & \cdot & \cdot & \cdot & \cdot & \cdot \\ \cdot & \cdot & \cdot & \cdot & \cdot & \cdot & \cdot \\ DO - Cd_i & DO - Ja_i & DO - Cu_i & DO - Pb_i & DO - Pl_i & E_i & Di \end{bmatrix}, i = 179$$